Statement of Compliance

Manufacturer:	Triangle Biomedical Sciences 3014 Crossdaile Dr Durham, NC 27705-2570 919-384-9393
Requester / Applicant:	Mario Moreno
Name of Equipment:	Automatic Tissue Processor
Model No.	ATP1-220
Type of Equipment:	Measurement, Control and Laboratory Use
Class of Equipment:	Class A
Application of Regulations:	EMC Directive 89/336/EEC
Test Dates:	1 January 2006 to 2 March 2006

Guidance Documents:

Emissions: EN61326:1997+A1:1998+A2:2001+A3:2003

Immunity: EN61326:1997+A1:1998+A2:2001+A3:2003

Test Methods:

Emissions: EN55022:1998+A1:2000+A2:2003, EN61000-3-2:2000, EN61000-3-3:1995+A1:2001

Immunity: EN61000-4-2:1995+A1:1998+A2:2001, EN61000-4-3:2002, EN61000-4-4:1995, EN61000-4-5:1995+A1:1996, EN61000-4-6:1996+A1:2001, EN61000-4-8:1995, EN61000-4-11:2001

The electromagnetic compatibility test and documented data described in this report has been performed and recorded by TUV Rheinland of North America, in accordance with the standards and procedures listed herein. As the responsible authorized agent of the EMC laboratory, I hereby declare that a sample of one, of the equipment described above, has been shown to be compliant with the EMC requirements of the stated regulations and standards based on these results. If any special accessories and/or modifications were required for compliance, they are listed in the Executive Summary of this report.

This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report contains data that are not covered by NVLAP accreditation. This report shall not be reproduced except in full, without the written authorization of the laboratory.

3 March 2006		3 March 2006	
Test Engineer	Date	NVLAP Signatory	Date

NVLAP

200094-0