



# **Declaration of Conformity**

**Company Name:** Triangle Biomedical Sciences, Inc.  
**Address:** 3014 Croasdaile Drive  
Durham, North Carolina 27705-2507  
U.S.A.

**Manufacturer:** Triangle Biomedical Sciences, Inc.  
**Address:** 3014 Croasdaile Drive  
Durham, North Carolina 27705-2507  
U.S.A.

**Product:** SMRGC-2 –**SHUR/Mount™** Robotic Glass Coverslipper

*The above mentioned products comply with the following European Directives:*

<b>Directive</b>	<b>Description</b>
LVD 2006/95/EC	Low Voltage Directive (LVD)
EMC 2004/108/EC	Electromagnetic Compatibility Directive (EMC)


*The conformity of the above mentioned product(s) to the Directives is supported by testing to the following Standards:*

<b>Standard</b>	<b>Description</b>
UL - 61010A CSA - 61010-1	Safety requirements for Electrical Equipment for Measurement, Control and Laboratory Use
CE - EN/IEC EN61010-1	Safety
CE - EN61326	Immunity
CE - EN55011	Emissions
CE - EN61000-3-3	Voltage Fluctuation / Flicker
CE - EN61000-3-2	Harmonic Distortion

**CE Mark applied:** January 2011

*I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s).*

**Signature of Representative  
for the manufacturer:**

Ken Marzinsky   
President, Triangle Biomedical Sciences, Inc. Date: January 1, 2011