



Declaration of Conformity

Company Name: Triangle Biomedical Sciences, Inc.
Address: 3014 Croasdaile Drive
Durham, North Carolina 27705-2507
U.S.A.

Manufacturer: Triangle Biomedical Sciences, Inc.
Address: 3014 Croasdaile Drive
Durham, North Carolina 27705-2507
U.S.A.

Product: TFB-120 and/or TFB-220 – Tissue Flotation Bath
TFB-L-120 and/or TFB-L-220 – Tissue Flotation Bath
TFB-S-120 and/or TFB-S-220 – Tissue Flotation Bath

FWS-120 and/or FWS-220 – Flotation Work Station
FWS-L-120 and/or FWS-L-220 – Flotation Work Station
FWS-S-120 and/or FWS-S-220 – Flotation Work Station

The above mentioned products comply with the following European Directives:

Directive	Description
LVD 2006/95/EC	Low Voltage Directive (LVD)
EMC 2004/108/EC	Electromagnetic Compatibility Directive (EMC)

The conformity of the above mentioned product(s) to the Directives is supported by testing to the following Standards:

Standard	Description
UL - 61010A CSA - 61010-1	Safety requirements for Electrical Equipment for Measurement, Control and Laboratory Use
CE - EN/IEC EN61010-1	Safety
CE - EN61326	Immunity
CE - EN55011	Emissions
CE - EN61000-3-3	Voltage Fluctuation / Flicker
CE - EN61000-3-2	Harmonic Distortion

CE Mark applied: January 2011

I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s).

**Signature of Representative
for the manufacturer:**

Ken Marzinsky 
President, Triangle Biomedical Sciences, Inc. Date: January 1, 2011