

Certificate



Certificate no.

CU 72052715 01

License Holder:

Triangle Biomedical Sciences
3014 Croasdaile Drive
Durham NC 27705
USA

Manufacturing Plant:

Triangle Biomedical Sciences
3014 Croasdaile Drive
Durham NC 27705
USA

Test report no.: USA-KK 30482592 001

Client Reference: Mario Moreno

Tested to: UL 61010-1:2004
CAN/CSA-C22.2 61010-1:2004

Certified Product: Laboratory Tissue Processing Equipment

License Fee - Units

Model Designation:

Paraffin Dispenser: 1) PD-120; 2) PD-220 7
Flotation Work Station: 3) FWS-120; 4) FWS-240
Tissue Flotation Bath: 5) TFB-120; 6) TFB-240

Rated Voltage: 1), 3), 5) AC 100-120V, 50/60Hz
2), 4), 6) AC 220-240V, 50/60Hz

Rated Current: 1) 4.5A; 2) 2.25A; 3) 5.0A
4) 2.5A; 5) 3.5A; 6) 1.75A

Protection Class: I

7

Appendix: 1, 1-5

Licensed Test mark:



Signatures

Date of Issue

(day/mo/yr)

21/10/2005

Stephan Schmitt
President

Dipl.-Ing. M. Raap
QA Certification Officer