



Certificate

Certificate no.

CU 72091145 01

License Holder:

Triangle Biomedical Sciences
3014 Croasdaile Drive
Durham NC 27705
USA

Manufacturing Plant:

Triangle Biomedical Sciences
3014 Croasdaile Drive
Durham NC 27705
USA

Test report no.: USA-JV 30981524 002

Client Reference: Jesse Hart

Tested to: UL 61010-1:2004 R7.05
CAN/CSA-C22.2 NO. 61010-1-04

Certified Product: Paraffin Block Trimmer

License Fee - Units

Model Designation: TBS Shur/Trim ST-120

7

Rated Voltage: AC 230-240V, 50/60Hz
Rated Current: 1.0A
Protection Class: I

Appendix: 1, 1-2

7

Inh. = 746031 / Deb. = 746031 / Fert. = 746031

Licensed Test mark:



Signature

Dipl.-Ing. M. Glagla
QA Certification Officer

Date of Issue

(day/mo/yr)

10/09/2009