



Certificate

Certificate no.

CU 72092645 02

License Holder:
Triangle Biomedical Sciences
3014 Croasdaile Drive
Durham NC 27705
USA

Manufacturing Plant:
Triangle Biomedical Sciences
3014 Croasdaile Drive
Durham NC 27705
USA

Test report no.: USA-SM 30982292 004 **Client Reference:** Eddie Hunnell

Tested to: UL 61010-1:2004 R10.08
CAN/CSA-C22.2 NO. 61010-1-04+GI1 (R2009)

Certified Product: Water Bath

License Fee - Units

Change:
Upgrade of standard for all models: see above

Addition:
Model Designation: 1) FWSS, 2) FWSL, 3) TFBS, 4) TFBL

Rated Current: 1) 3.7A; 2) 5.4A; 3) 3.4A; 4) 5.0A

Appendix: 1, 1-6

Inh. = 746031 / Deb. = 746031 / Fert. = 746031

Licensed Test mark:



Signature

Dipl.-Ing. M. Glagla
QA Certification Officer

Date of Issue

(day/mo/yr)

20/08/2010